



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/891,572	06/25/2001	Alan M. Cohen	BMT-009AUS	1952
22494	7590	02/27/2004	EXAMINER	
DALY, CROWLEY & MOFFORD, LLP			ALEXANDER, LYLE	
SUITE 101			ART UNIT	
275 TURNPIKE STREET			PAPER NUMBER	
CANTON, MA 02021-2310			1743	

DATE MAILED: 02/27/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/891,572

Applicant(s)

COHEN ET AL.

Examiner

Lyle A Alexander

Art Unit

1743

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. ____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date ____.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date ____.
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: ____.

Art Unit: 1743

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1,5-6 and 17-20 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Heinonen et al.

Heinonen et al. teach glucose monitoring apparatus that calculates glucose blood levels, remotely transmits the data to the physician and provides a corrector(13) to continuously change the correction coefficients used in the calculations. A measurement device(1) having wireless data transmission capabilities transmits the measurements to monitoring equipment(9) that will calculate the suggested medication, diet and exercise levels for the patient. Additionally, the corrector(13) calculates a predictive glucose value that has been read on the claimed "permissible glucose deviation". The corrector(13) has been read on claimed processing of the glucose measurement. With respect to the limitation that the patient is identified, it would have been inherent that the transmission to the physician would have contained the appropriate information to identify the patient thus enabling the physician to make a diagnosis.

Claims 1-3,5-10 and 17-20 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Allen.

Art Unit: 1743

Allen (USP 4,731,726) teaches a blood glucose monitoring system, using a convention blood glucose test strip that transmits the measured glucose values to the health care provider. Column 1 lines 50 through column 2 describe a physician prescribed algorithm that accounts for insulin administration data, stress, exercise and dietary intake. Subsequent information is sent back to the patient in response to the physician's algorithm concerning medications. This step has been read on the claimed ~~teach the glucose values~~ processing of the glucose measurements. Column 5 lines 24-36 teach the phone modem and the use of a glucose test strip. With respect to the limitation that the patient is identified, it would have been inherent that the transmission to the physician would have contained the appropriate information to identify the patient enabling the physician to make a diagnosis.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Art Unit: 1743

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 4 and 11-16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Allen in view of Tajnafoi (UPS 6,48,067).

See Allen *supra*.

Allen is silent to the claimed compensating for the aging of the test strip.

Tajnafoi teaches in column 6 lines 4+ teaches an algorithm that interpolates the glucose values from the sample which aids in minimizing errors from the aging of the test strip, ambient temperature, vapor content, etc.

It would have been within the skill of the art to modify Allen in view of Tajnafoi and compensate for the age of the test strip to minimize any errors that could result in an improper dosage of insulin.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Lyle A Alexander whose telephone number is 571-272-1254. The examiner can normally be reached on Monday, Wednesday and Friday.

Art Unit: 1743

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill Warden can be reached on 571-272-1267. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Lyle A Alexander
Primary Examiner
Art Unit 1743
